

# Test Report

**Applicant:** Ningbo Haishu Junheng Fashion Co., Ltd.  
**Address:** Fengjia Village shiqi Haishu Areh Ningbo Zhejiang China

**The following sample(s) was/were submitted and identified on behalf of the client as:**

Product name: Protective Mask  
Model: JH-032  
Trade mark: JH  
Manufacturer: Ningbo Haishu Junheng Fashion Co., Ltd.  
Address: Fengjia Village shiqi Haishu Areh Ningbo Zhejiang China  
Classification: Type IIR  
Sample quantity: 58 Pcs  
Sample Received Date: Nov. 09, 2020  
Testing Period: Nov. 09, 2020~ Nov. 12, 2020

**Test Requirement:**

According to the requirement of the client, the test item of the sample is according to the standard EN 14683:2019+AC:2019.

**Test Result(s):** Please refer to the following page(s)

**Test Method:** Please refer to the following page(s)

Compiled by: Nancy

Reviewed by: ADa

Approved by: Mark Liaw, Tang Jie

Date: 2020-11-12

**Test Result(s):**

Tests performed	Test Standards	Test Requirements	Test results		Conclusion
1. Bacterial filtration efficiency (BFE)*	EN 14683-2019 Clause 5.2.2	Bacterial filtration efficiency:  The BFE shall conform to the minimum value given for the relevant type when tested in accordance Annex B:  Type I: ≥95%, Type II & IIR: ≥98%	1#	99.74%	Pass
			2#	99.96%	
			3#	99.89%	
			4#	99.85%	
			5#	99.85%	
2. Splash resistance	EN 14683-2019 Clause 5.2.4	Splash resistance:  The splash resistance shall conform to the minimum value for type IIR when tested in accordance with ISO 22609:2004: ≥ 16.0 kPa	The specimens have been tested without penetration by Synthetic Blood		Pass

Remark:

\*Each test specimen has been conditioned at (21 ± 5) °C and (85 ± 5) % relative humidity for 4 h.

**Sample photo(s):**



Fig.1



Fig.2



Fig.3

\*\*\*\*End of Report\*\*\*\*

The test report is effective only with both signature and specialized stamp, the result(s) shown in this report refer only to the sample(s) tested. Without written approval of NTEK, this report can't be reproduced except in full;The laboratory is not responsible for the authenticity of the sample information provided by the customer;The laboratory is not responsible for any deviation of results due to methods/standards provided by the customer.